
Midmark M11 - v1.0.0
Total Cycles: 3718

Name: RITTER 1
Sterilizer ID: V2178198
09/05/2023 12:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.5 F 0.0
2:00 151.7 F 0.0
4:00 176.3 F 2.6
6:00 200.5 F 6.4
8:00 218.4 F 11.0
10:00 224.0 F 6.1
12:00 235.8 F 11.1
14:00 233.0 F 8.0
16:00 242.6 F 11.5
18:00 245.5 F 12.9
20:00 257.9 F 19.5
22:00 269.4 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.4 F 28.7
2:00 271.5 F 28.3
3:00 272.5 F 28.8
4:00 271.5 F 28.1

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.5 F 0.0
5:00 216.2 F 0.0
10:00 220.1 F 0.0
15:00 233.3 F 0.0
20:00 228.7 F 0.0
25:00 230.2 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 23:02
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:01:11
CYCLE COMPLETE

09/05/2023 01:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____