
Midmark M11 - v1.0.5
Total Cycles: 1865

Name: RITTER2
Sterilizer ID: V2313228
12/20/2022 03:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.9 F 0.0
2:00 196.8 F 4.3
4:00 207.1 F 9.8
6:00 217.8 F 6.2
8:00 231.8 F 11.3
10:00 235.0 F 9.2
12:00 238.0 F 9.4
14:00 250.1 F 15.6
16:00 262.9 F 23.1
18:00 271.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.3 F 29.3
2:00 271.7 F 29.3
3:00 272.2 F 29.3
4:00 271.5 F 29.1

Min 271.4 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 225.0 F 0.0
10:00 231.6 F 0.0
15:00 229.9 F 0.0
20:00 227.4 F 0.0
25:00 229.2 F 0.0
30:00 230.0 F 0.0
Drying Complete

FILLING: 2:02
HEATING: 18:04
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:18
CYCLE COMPLETE

12/20/2022 04:03 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____