

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4920

Name: RITTER 1  
Sterilizer ID: V2178198  
09/19/2024 09:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.2 F 0.0  
2:00 70.5 F 0.0  
4:00 79.9 F 0.0  
6:00 115.1 F 0.1  
8:00 145.4 F 0.2  
10:00 166.3 F 0.3  
12:00 189.3 F 1.2  
14:00 209.7 F 4.2  
16:00 213.4 F 7.6  
18:00 224.1 F 11.0  
20:00 228.3 F 6.5  
22:00 235.7 F 10.6  
24:00 235.3 F 8.0  
26:00 246.5 F 13.5  
28:00 244.4 F 12.2  
30:00 256.6 F 18.6  
32:00 266.9 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.3 F 28.7  
2:00 272.6 F 28.8  
3:00 272.1 F 28.3  
4:00 271.5 F 28.1  
  
Min 271.5 F 28.0  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.3 F 0.0  
5:00 201.5 F 0.0  
10:00 204.1 F 0.0  
15:00 207.2 F 0.0  
20:00 211.3 F 0.0  
25:00 221.5 F 0.0  
30:00 230.4 F 0.0  
  
Drying Complete

FILLING: 0:42  
HEATING: 33:41  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:11:56  
CYCLE COMPLETE

09/19/2024 10:13 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_