
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003853
13:22 21/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:30
STERILIZING 3:30
136.0 C 220kPa 3:30
136.2 C 221kPa 4:00
136.4 C 222kPa 4:30
136.4 C 221kPa 5:00
135.8 C 219kPa 5:30
136.2 C 223kPa 6:00
136.2 C 224kPa 6:30
136.0 C 217kPa 7:00
136.0 C 217kPa 7:00
VENTING 7:00
AIR DRYING 8:47
CYCLE COMPLETE 9:16

Digital Signature #
BD06B4449F5C9C67
