
Midmark M11 - v1.0.5
Total Cycles: 222

Name: RITTER2
Sterilizer ID: V2313228
07/19/2021 10:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.5 F 0.0
2:00 178.8 F 2.7
4:00 179.8 F 6.9
6:00 197.2 F 11.9
8:00 216.3 F 7.6
10:00 226.7 F 13.0
12:00 236.6 F 10.9
14:00 236.1 F 9.4
16:00 249.7 F 15.9
18:00 261.8 F 22.7
20:00 272.7 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 30.2
1:00 272.3 F 30.2
2:00 272.4 F 30.3
3:00 272.3 F 30.5
4:00 272.1 F 30.3

Min 271.8 F 30.2
Max 272.6 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.6 F 0.0
5:00 238.3 F 0.0
10:00 229.5 F 0.0
15:00 221.4 F 0.0
20:00 219.5 F 0.0
25:00 176.2 F 0.0
30:00 169.0 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 20:58
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:47
CYCLE COMPLETE

07/19/2021 11:20 AM
ACCEPTED: X
REJECTED:
Operator: U01