

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 007215  
12:34 17/04/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:22  
PRESSURIZING 2:25  
STERILIZING 3:20  
136.3 C 222kPa 3:20  
136.3 C 222kPa 3:50  
136.2 C 224kPa 4:20  
136.1 C 221kPa 4:50  
136.4 C 222kPa 5:20  
136.4 C 224kPa 5:50  
136.4 C 223kPa 6:20  
136.4 C 224kPa 6:50  
136.4 C 224kPa 6:50  
VENTING 6:50  
AIR DRYING 7:39  
CYCLE COMPLETE 21:29

-----

Digital Signature #  
C5FC3774DC1BD508  
-----