

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5180

Name: RITTER 1  
Sterilizer ID: V2178198  
12/11/2024 03:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 98.2 F 0.0  
2:00 100.1 F 0.0  
4:00 110.8 F 0.0  
6:00 150.6 F 0.1  
8:00 177.7 F 0.3  
10:00 204.5 F 0.8  
12:00 222.6 F 4.1  
14:00 230.0 F 7.0  
16:00 238.2 F 10.4  
18:00 234.2 F 8.9  
20:00 241.5 F 11.2  
22:00 235.3 F 8.5  
24:00 242.5 F 11.3  
26:00 246.2 F 13.2  
28:00 257.1 F 19.0  
30:00 267.0 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.0 F 28.6  
2:00 272.5 F 29.0  
3:00 272.7 F 28.9  
4:00 272.4 F 28.7

Min 271.4 F 28.1  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.0 F 0.0  
5:00 244.6 F 0.0  
10:00 229.0 F 0.0  
15:00 226.8 F 0.0  
20:00 229.5 F 0.0  
25:00 229.3 F 0.0  
30:00 222.1 F 0.0  
Drying Complete

FILLING: 0:44  
HEATING: 31:49  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:10:25  
CYCLE COMPLETE

12/11/2024 04:18 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_