
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000821
14:32 13/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:48
PRESSURIZING 2:58
STERILIZING 4:14
136.0 C 228kPa 4:14
136.3 C 231kPa 4:44
136.4 C 233kPa 5:14
136.5 C 233kPa 5:44
136.4 C 233kPa 6:14
136.4 C 232kPa 6:44
136.3 C 232kPa 7:14
136.2 C 231kPa 7:44
136.2 C 231kPa 7:44
VENTING 7:44
AIR DRYING 8:35
CYCLE COMPLETE 23:36

Digital Signature #
8653B3B07EA54F83
