
Midmark M11 - v1.0.5
Total Cycles: 3900

Name: RITTER2
Sterilizer ID: V2313228
08/28/2024 01:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 92.6 F 0.0
2:00 92.9 F 0.0
4:00 97.0 F 0.2
6:00 138.6 F 1.2
8:00 171.9 F 4.3
10:00 199.8 F 7.7
12:00 208.8 F 11.6
14:00 218.3 F 5.2
16:00 221.0 F 9.5
18:00 236.1 F 11.1
20:00 236.5 F 11.0
22:00 235.5 F 9.0
24:00 248.1 F 15.0
26:00 258.6 F 21.0
28:00 268.1 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.6
1:00 272.9 F 30.3
2:00 271.8 F 29.5
3:00 272.7 F 30.4
4:00 272.4 F 29.9

Min 271.4 F 29.5
Max 273.0 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 216.4 F 0.0
10:00 210.2 F 0.0
15:00 207.0 F 0.0
20:00 205.2 F 0.0
25:00 213.4 F 0.0
30:00 221.7 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 29:21
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:07:51
CYCLE COMPLETE

08/28/2024 03:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____