

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2472

Name: RITTER2  
Sterilizer ID: V2313228  
06/06/2023 12:14 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.5 F 0.0  
2:00 179.4 F 4.1  
4:00 197.4 F 9.3  
6:00 219.1 F 7.8  
8:00 229.9 F 10.5  
10:00 231.4 F 7.6  
12:00 241.1 F 11.5  
14:00 245.3 F 13.5  
16:00 257.9 F 20.3  
18:00 269.0 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.2  
1:00 272.5 F 29.7  
2:00 271.3 F 29.1  
3:00 272.6 F 29.8  
4:00 271.6 F 29.1  
  
Min 271.2 F 29.0  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.8 F 0.0  
5:00 213.5 F 0.0  
10:00 232.5 F 0.0  
15:00 229.8 F 0.0  
20:00 227.0 F 0.0  
25:00 224.9 F 0.0  
30:00 223.0 F 0.0  
Drying Complete

FILLING: 1:56  
HEATING: 19:06  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:18  
CYCLE COMPLETE

06/06/2023 01:12 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_