

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 007397  
6:55 21/05/2024

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:48  
PRESSURIZING 3:49  
STERILIZING 4:51  
136.4 C 222kPa 4:51  
136.4 C 224kPa 5:21  
136.4 C 221kPa 5:51  
136.2 C 222kPa 6:21  
136.4 C 220kPa 6:51  
136.2 C 219kPa 7:21  
136.3 C 221kPa 7:51  
136.4 C 221kPa 8:21  
136.4 C 221kPa 8:21  
VENTING 8:21  
AIR DRYING 9:11  
CYCLE COMPLETE 39:11

-----

Digital Signature #  
688F3D93E5467C1E  
-----