
Midmark M11 - v1.0.0
Total Cycles: 3671

Name: RITTER 1
Sterilizer ID: V2178198
08/17/2023 12:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.6 F 0.0
2:00 151.9 F 2.8
4:00 169.5 F 6.7
6:00 193.9 F 11.5
8:00 212.5 F 5.0
10:00 227.1 F 9.8
12:00 231.7 F 8.5
14:00 240.1 F 11.7
16:00 238.5 F 9.8
18:00 251.7 F 16.3
20:00 263.1 F 22.9
22:00 272.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.3 F 28.8
2:00 272.6 F 29.3
3:00 271.7 F 28.4
4:00 272.7 F 29.2

Min 271.5 F 28.4
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.4 F 0.0
5:00 224.3 F 0.0
10:00 230.8 F 0.0
15:00 228.5 F 0.0
20:00 227.8 F 0.0
25:00 229.0 F 0.0
30:00 230.9 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 22:11
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:34
CYCLE COMPLETE

08/17/2023 01:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____