

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000094  
7:25 02/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:27  
PRESSURIZING 2:33  
STERILIZING 3:47  
136.4 C 224kPa 3:47  
136.4 C 225kPa 4:17  
136.4 C 223kPa 4:47  
136.3 C 224kPa 5:17  
136.4 C 224kPa 5:47  
136.3 C 223kPa 6:17  
136.4 C 224kPa 6:47  
136.3 C 224kPa 7:17  
136.3 C 224kPa 7:17  
VENTING 7:17  
AIR DRYING 8:07  
CYCLE COMPLETE 8:27  
-----

Digital Signature #  
DDAA5C87289D5EE6  
-----