
Midmark M11 - v1.0.0
Total Cycles: 1185

Name: RITTER 1
Sterilizer ID: V2178198
09/20/2021 08:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.6 F 0.0
2:00 159.6 F 1.2
4:00 166.0 F 3.7
6:00 181.8 F 6.8
8:00 199.5 F 10.9
10:00 217.2 F 5.0
12:00 227.8 F 9.1
14:00 236.4 F 11.1
16:00 239.3 F 11.1
18:00 236.6 F 9.0
20:00 250.2 F 15.5
22:00 261.2 F 21.8
24:00 271.6 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.5
1:00 272.1 F 29.0
2:00 272.5 F 29.0
3:00 271.5 F 28.4
4:00 272.3 F 29.2

Min 271.5 F 28.4
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 207.5 F 0.0
10:00 197.4 F 0.0
15:00 192.8 F 0.0
20:00 194.7 F 0.0
25:00 192.7 F 0.0
30:00 200.0 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 24:31
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:02:55
CYCLE COMPLETE

09/20/2021 09:19 AM
ACCEPTED: X
REJECTED:
Operator: U01