
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006297
14:56 12/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:05
STERILIZING 4:14
136.4 C 220kPa 4:14
136.4 C 223kPa 4:44
136.4 C 224kPa 5:14
136.4 C 224kPa 5:44
136.4 C 223kPa 6:14
136.4 C 225kPa 6:44
136.4 C 224kPa 7:14
136.4 C 225kPa 7:44
136.4 C 225kPa 7:44
VENTING 7:44
AIR DRYING 8:39
CYCLE COMPLETE 13:48

Digital Signature #
B0080E3182F07B6D
