
Midmark M11 - v1.0.0
Total Cycles: 1709

Name: RITTER 1
Sterilizer ID: V2178198
03/14/2022 07:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.8 F 0.0
2:00 72.4 F 0.0
4:00 80.7 F 0.4
6:00 114.9 F 1.5
8:00 146.5 F 2.6
10:00 171.7 F 4.3
12:00 189.9 F 7.0
14:00 204.5 F 10.4
16:00 212.8 F 7.2
18:00 225.8 F 6.9
20:00 234.0 F 11.0
22:00 232.2 F 7.1
24:00 244.4 F 13.2
26:00 242.6 F 11.5
28:00 254.7 F 17.8
30:00 265.3 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.5 F 28.9
2:00 271.4 F 28.2
3:00 272.7 F 29.0
4:00 271.6 F 28.2

Min 271.3 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 172.3 F 0.0
5:00 213.0 F 0.0
10:00 223.1 F 0.0
15:00 221.2 F 0.0
20:00 236.7 F 0.0
25:00 226.7 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:52
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:09:42
CYCLE COMPLETE

03/14/2022 08:12 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____