
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006096
10:41 10/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:26
STERILIZING 3:29
136.4 C 224kPa 3:29
136.2 C 223kPa 3:59
136.1 C 222kPa 4:29
136.4 C 224kPa 4:59
136.3 C 224kPa 5:29
136.0 C 222kPa 5:59
136.2 C 223kPa 6:29
136.4 C 223kPa 6:59
136.4 C 223kPa 6:59
VENTING 6:59
AIR DRYING 7:49
CYCLE COMPLETE 22:37

Digital Signature #
94C2F9B036B26D19
