
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000980
10:18 10/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:14
PRESSURIZING 2:21
STERILIZING 3:24
136.0 C 227kPa 3:24
136.0 C 230kPa 3:54
136.4 C 233kPa 4:24
136.3 C 232kPa 4:54
136.4 C 233kPa 5:24
136.4 C 232kPa 5:54
136.4 C 233kPa 6:24
136.4 C 233kPa 6:54
136.4 C 233kPa 6:54
VENTING 6:54
AIR DRYING 7:46
CYCLE COMPLETE 8:09

Digital Signature #
9EE846D4A16627C9
