
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 005188
17:19 01/03/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:28
STERILIZING 3:24
136.4 C 220kPa 3:24
136.2 C 220kPa 3:54
136.4 C 224kPa 4:24
136.3 C 223kPa 4:54
136.3 C 224kPa 5:24
136.4 C 222kPa 5:54
136.5 C 225kPa 6:24
136.4 C 220kPa 6:54
136.4 C 220kPa 6:54
VENTING 6:55
AIR DRYING 7:47
CYCLE COMPLETE 16:21

Digital Signature #
83BAE3DAC39E06E2
