

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2489

Name: RITTER 1  
Sterilizer ID: V2178198  
10/26/2022 07:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.8 F 0.0  
2:00 74.2 F 0.0  
4:00 84.7 F 0.0  
6:00 115.3 F 1.4  
8:00 140.9 F 3.9  
10:00 166.1 F 7.0  
12:00 183.9 F 10.2  
14:00 200.2 F 11.6  
16:00 210.2 F 4.7  
18:00 220.2 F 8.0  
20:00 229.7 F 11.6  
22:00 230.4 F 6.7  
24:00 239.6 F 11.8  
26:00 237.0 F 9.0  
28:00 249.2 F 14.8  
30:00 259.0 F 20.4  
32:00 268.3 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.5 F 29.3  
2:00 271.7 F 28.6  
3:00 271.6 F 28.8  
4:00 272.4 F 29.2  
  
Min 271.2 F 28.3  
Max 272.5 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.1 F 0.0  
5:00 201.8 F 0.0  
10:00 210.0 F 0.0  
15:00 209.6 F 0.0  
20:00 218.9 F 0.0  
25:00 223.0 F 0.0  
30:00 231.1 F 0.0  
  
Drying Complete

FILLING: 0:45  
HEATING: 33:23  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:11:20  
CYCLE COMPLETE

10/26/2022 09:00 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_