
Midmark M11 - v1.0.5
Total Cycles: 514

Name: RITTER2
Sterilizer ID: V2313228
11/02/2021 11:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.2 F 0.0
2:00 176.6 F 2.9
4:00 183.0 F 7.5
6:00 203.2 F 12.5
8:00 220.1 F 7.1
10:00 233.2 F 12.3
12:00 236.2 F 10.3
14:00 236.8 F 9.0
16:00 250.0 F 16.0
18:00 261.6 F 22.6
20:00 272.1 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.2
1:00 272.7 F 30.0
2:00 271.8 F 29.1
3:00 272.7 F 30.0
4:00 272.0 F 29.2

Min 271.7 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.2 F 0.0
5:00 189.4 F 0.0
10:00 191.6 F 0.0
15:00 193.2 F 0.0
20:00 193.0 F 0.0
25:00 192.8 F 0.0
30:00 191.5 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 20:22
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

11/02/2021 12:32 PM
ACCEPTED: X
REJECTED:
Operator: U01