
Midmark M11 - v1.0.0
Total Cycles: 4731

Name: RITTER 1
Sterilizer ID: V2178198
07/15/2024 03:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.3 F 0.0
2:00 173.4 F 3.4
4:00 182.7 F 7.5
6:00 204.8 F 12.6
8:00 219.9 F 6.5
10:00 230.0 F 12.0
12:00 233.6 F 9.1
14:00 237.5 F 9.6
16:00 247.2 F 14.2
18:00 259.1 F 20.8
20:00 270.1 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.6 F 29.3
2:00 271.2 F 28.9
3:00 272.6 F 29.3
4:00 271.1 F 28.7

Min 271.0 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.9 F 0.0
5:00 215.1 F 0.0
10:00 223.7 F 0.0
15:00 228.1 F 0.0
20:00 228.0 F 0.0
25:00 228.4 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 20:51
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 00:59:56
CYCLE COMPLETE

07/15/2024 04:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____