
Midmark M11 - v1.0.5
Total Cycles: 3727

Name: RITTER2
Sterilizer ID: V2313228
06/21/2024 11:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.8 F 0.0
2:00 153.5 F 2.2
4:00 166.6 F 5.7
6:00 191.6 F 10.3
8:00 212.9 F 5.2
10:00 225.1 F 9.7
12:00 231.4 F 8.6
14:00 241.1 F 13.0
16:00 241.8 F 12.0
18:00 254.8 F 18.8
20:00 266.4 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.4
1:00 272.8 F 30.1
2:00 271.9 F 29.4
3:00 272.0 F 30.0
4:00 272.8 F 30.0

Min 271.4 F 29.2
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.8 F 0.0
5:00 203.7 F 0.0
10:00 205.7 F 0.0
15:00 203.2 F 0.0
20:00 203.8 F 0.0
25:00 201.5 F 0.0
30:00 205.2 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 21:37
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:00:19
CYCLE COMPLETE

06/21/2024 12:37 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____