
Midmark M11 - v1.0.5
Total Cycles: 3792

Name: RITTER2
Sterilizer ID: V2313228
07/17/2024 09:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.5 F 0.0
2:00 72.7 F 0.0
4:00 76.7 F 0.0
6:00 88.7 F 0.1
8:00 121.7 F 0.1
10:00 155.5 F 1.4
12:00 178.6 F 4.2
14:00 203.3 F 7.1
16:00 215.4 F 10.3
18:00 224.9 F 10.3
20:00 227.5 F 8.6
22:00 237.5 F 12.9
24:00 236.1 F 10.0
26:00 237.0 F 9.5
28:00 246.0 F 13.9
30:00 256.4 F 19.5
32:00 266.0 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 272.4 F 30.0
2:00 272.1 F 29.4
3:00 271.1 F 29.1
4:00 272.0 F 30.0

Min 271.1 F 29.0
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.5 F 0.0
5:00 203.8 F 0.0
10:00 222.6 F 0.0
15:00 217.1 F 0.0
20:00 223.3 F 0.0
25:00 228.0 F 0.0
30:00 226.0 F 0.0

Drying Complete

FILLING: 0:55
HEATING: 33:53
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:12:33
CYCLE COMPLETE

07/17/2024 10:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____