
Midmark M11 - v1.0.5
Total Cycles: 1154

Name: RITTER2
Sterilizer ID: V2313228
06/01/2022 03:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.8 F 0.0
2:00 156.0 F 2.6
4:00 173.6 F 6.2
6:00 196.2 F 11.0
8:00 218.7 F 5.2
10:00 225.9 F 10.5
12:00 229.5 F 7.6
14:00 244.4 F 13.4
16:00 242.9 F 12.1
18:00 255.6 F 18.7
20:00 266.8 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.5 F 29.6
2:00 272.5 F 29.5
3:00 273.0 F 29.8
4:00 272.5 F 30.1

Min 271.0 F 28.8
Max 273.4 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.7 F 0.0
5:00 208.1 F 0.0
10:00 225.7 F 0.0
15:00 224.8 F 0.0
20:00 225.2 F 0.0
25:00 225.6 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 21:31
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:47
CYCLE COMPLETE

06/01/2022 04:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____