
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001060
10:34 06/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:47
PRESSURIZING 2:55
STERILIZING 4:05
136.4 C 220kPa 4:05
136.4 C 223kPa 4:35
136.4 C 225kPa 5:05
136.3 C 224kPa 5:35
136.4 C 224kPa 6:05
136.0 C 222kPa 6:35
136.4 C 223kPa 7:05
136.1 C 222kPa 7:35
136.1 C 222kPa 7:35
VENTING 7:35
AIR DRYING 8:24
CYCLE COMPLETE 23:24

Digital Signature #
7649F323E9E31AE4
