
Midmark M11 - v1.0.0
Total Cycles: 2759

Name: RITTER 1
Sterilizer ID: V2178198
01/10/2023 02:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.6 F 0.0
2:00 138.7 F 0.0
4:00 168.8 F 0.8
6:00 193.7 F 5.2
8:00 210.9 F 9.7
10:00 225.0 F 8.0
12:00 232.7 F 9.9
14:00 235.7 F 8.7
16:00 245.1 F 13.1
18:00 243.3 F 11.7
20:00 256.0 F 18.4
22:00 267.2 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.2 F 28.8
2:00 271.3 F 28.2
3:00 272.3 F 28.9
4:00 271.5 F 28.2

Min 271.2 F 28.1
Max 272.4 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.3 F 0.0
5:00 215.7 F 0.0
10:00 224.2 F 0.0
15:00 229.1 F 0.0
20:00 232.1 F 0.0
25:00 227.5 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 23:28
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:31
CYCLE COMPLETE

01/10/2023 03:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____