
Midmark M11 - v1.0.0
Total Cycles: 3196

Name: RITTER 1
Sterilizer ID: V2178198
05/03/2023 01:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.0 F 0.0
2:00 152.0 F 0.0
4:00 175.5 F 1.9
6:00 200.1 F 6.1
8:00 216.9 F 11.2
10:00 224.2 F 6.4
12:00 236.2 F 11.8
14:00 234.9 F 9.1
16:00 238.6 F 9.5
18:00 247.6 F 14.3
20:00 260.0 F 21.2
22:00 271.2 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.6 F 29.2
2:00 271.8 F 28.9
3:00 272.5 F 29.2
4:00 271.7 F 28.6

Min 271.6 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.2 F 0.0
5:00 206.0 F 0.0
10:00 227.3 F 0.0
15:00 229.0 F 0.0
20:00 229.0 F 0.0
25:00 229.9 F 0.0
30:00 230.0 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 22:40
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:00:50
CYCLE COMPLETE

05/03/2023 02:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____