
Midmark M11 - v1.0.5
Total Cycles: 2839

Name: RITTER2
Sterilizer ID: V2313228
09/06/2023 02:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.1 F 0.0
2:00 146.9 F 0.5
4:00 169.5 F 3.9
6:00 191.6 F 8.2
8:00 207.4 F 12.7
10:00 222.5 F 7.4
12:00 233.9 F 12.1
14:00 234.7 F 9.1
16:00 239.3 F 10.4
18:00 245.9 F 13.6
20:00 257.4 F 19.8
22:00 267.7 F 26.3
24:00 271.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.3 F 29.5
2:00 272.7 F 29.9
3:00 271.3 F 29.4
4:00 272.5 F 29.6

Min 271.2 F 29.2
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.9 F 0.0
5:00 224.2 F 0.0
10:00 191.9 F 0.0
15:00 189.3 F 0.0
20:00 186.1 F 0.0
25:00 184.0 F 0.0
30:00 182.3 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 24:06
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:02:34
CYCLE COMPLETE

09/06/2023 03:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____