
Midmark M11 - v1.0.5
Total Cycles: 939

Name: RITTER2
Sterilizer ID: V2313228
03/31/2022 01:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.0 F 0.0
2:00 196.5 F 3.7
4:00 209.9 F 9.8
6:00 224.1 F 7.3
8:00 237.7 F 13.2
10:00 239.9 F 12.0
12:00 241.1 F 11.7
14:00 255.3 F 19.2
16:00 267.8 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.6 F 30.1
2:00 271.5 F 29.3
3:00 272.5 F 30.3
4:00 272.4 F 29.8

Min 271.3 F 29.2
Max 272.8 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 222.9 F 0.0
10:00 238.1 F 0.0
15:00 233.4 F 0.0
20:00 226.0 F 0.0
25:00 223.1 F 0.0
30:00 224.2 F 0.0
Drying Complete

FILLING: 2:35
HEATING: 17:16
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:01
CYCLE COMPLETE

03/31/2022 02:24 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____