

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1191

Name: RITTER2  
Sterilizer ID: V2313228  
06/15/2022 12:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.2 F 0.0  
2:00 171.6 F 3.7  
4:00 187.0 F 8.1  
6:00 212.2 F 12.0  
8:00 226.0 F 8.8  
10:00 235.4 F 10.3  
12:00 241.9 F 12.6  
14:00 242.6 F 11.9  
16:00 255.7 F 18.8  
18:00 267.3 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 272.6 F 29.7  
2:00 271.7 F 29.0  
3:00 271.7 F 29.4  
4:00 272.6 F 29.7  
  
Min 271.2 F 28.8  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.2 F 0.0  
5:00 224.0 F 0.0  
10:00 226.2 F 0.0  
15:00 225.7 F 0.0  
20:00 225.0 F 0.0  
25:00 224.2 F 0.0  
30:00 223.3 F 0.0  
  
Drying Complete

FILLING: 1:35  
HEATING: 19:27  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:10  
CYCLE COMPLETE

06/15/2022 12:59 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_