
Midmark M11 - v1.0.0
Total Cycles: 1368

Name: RITTER 1
Sterilizer ID: V2178198
11/15/2021 10:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.7 F 0.0
2:00 178.5 F 2.8
4:00 190.7 F 7.0
6:00 218.6 F 12.7
8:00 226.4 F 7.5
10:00 240.5 F 12.6
12:00 237.9 F 9.8
14:00 236.3 F 8.5
16:00 248.8 F 14.7
18:00 260.7 F 21.3
20:00 271.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.8 F 29.1
2:00 271.6 F 28.3
3:00 272.7 F 29.2
4:00 272.2 F 28.7

Min 271.3 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.5 F 0.0
5:00 253.3 F 0.0
10:00 196.6 F 0.0
15:00 195.2 F 0.0
20:00 191.1 F 0.0
25:00 188.4 F 0.0
30:00 191.1 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 20:35
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:35
CYCLE COMPLETE

11/15/2021 11:22 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____