
Midmark M11 - v1.0.5
Total Cycles: 3050

Name: RITTER2
Sterilizer ID: V2313228
10/24/2023 01:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.6 F 0.0
2:00 124.5 F 1.0
4:00 145.5 F 3.9
6:00 176.2 F 8.2
8:00 198.7 F 13.4
10:00 227.1 F 7.5
12:00 230.7 F 12.5
14:00 234.9 F 9.2
16:00 238.1 F 9.6
18:00 247.0 F 14.0
20:00 259.2 F 20.8
22:00 270.8 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.4
1:00 272.0 F 29.3
2:00 272.7 F 29.5
3:00 272.4 F 29.8
4:00 272.5 F 30.0

Min 271.5 F 29.2
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.0 F 0.0
5:00 234.6 F 0.0
10:00 205.9 F 0.0
15:00 189.0 F 0.0
20:00 184.4 F 0.0
25:00 181.9 F 0.0
30:00 180.2 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 23:29
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:01:44
CYCLE COMPLETE

10/24/2023 02:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____