

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 007193  
7:54 12/04/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:02  
PRESSURIZING 2:07  
STERILIZING 3:09  
136.0 C 220kPa 3:09  
136.3 C 222kPa 3:39  
136.4 C 223kPa 4:09  
136.4 C 222kPa 4:39  
136.2 C 222kPa 5:09  
136.4 C 222kPa 5:39  
136.4 C 223kPa 6:09  
136.4 C 224kPa 6:39  
136.4 C 224kPa 6:39  
VENTING 6:39  
AIR DRYING 7:26  
CYCLE COMPLETE 37:26

-----  
Digital Signature #  
C31FC83A2BE8C83C  
-----