
Midmark M11 - v1.0.0
Total Cycles: 4226

Name: RITTER 1
Sterilizer ID: V2178198
01/24/2024 04:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.2 F 0.0
2:00 127.3 F 1.5
4:00 145.8 F 4.3
6:00 173.4 F 7.6
8:00 199.2 F 11.7
10:00 216.7 F 4.9
12:00 229.5 F 9.0
14:00 235.1 F 13.4
16:00 235.8 F 9.4
18:00 240.6 F 10.4
20:00 244.7 F 12.6
22:00 255.8 F 18.4
24:00 266.0 F 24.6
26:00 272.8 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.0 F 28.6
2:00 272.5 F 29.0
3:00 271.6 F 29.0
4:00 271.8 F 28.6

Min 271.2 F 28.5
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.7 F 0.0
5:00 207.3 F 0.0
10:00 252.2 F 0.0
15:00 246.0 F 0.0
20:00 220.4 F 0.0
25:00 215.0 F 0.0
30:00 208.0 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 26:22
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:05:22
CYCLE COMPLETE

01/24/2024 05:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____