
Midmark M11 - v1.0.0
Total Cycles: 896

Name: RITTER 1
Sterilizer ID: V2178198
06/15/2021 06:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.2 F 0.0
2:00 76.2 F 0.0
4:00 81.2 F 0.1
6:00 109.3 F 0.1
8:00 145.4 F 0.2
10:00 182.4 F 0.3
12:00 204.0 F 3.2
14:00 223.4 F 6.4
16:00 231.3 F 9.7
18:00 239.7 F 12.3
20:00 238.3 F 10.1
22:00 235.3 F 8.3
24:00 245.9 F 12.7
26:00 244.9 F 12.5
28:00 257.7 F 19.2
30:00 268.9 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.2 F 28.2
2:00 272.6 F 28.8
3:00 271.4 F 28.2
4:00 272.2 F 28.3

Min 271.3 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.8 F 0.0
5:00 226.4 F 0.0
10:00 229.9 F 0.0
15:00 230.9 F 0.0
20:00 225.9 F 0.0
25:00 228.9 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 31:04
STERILIZING: 4:00
VENTING: 3:00
DRYING: 30:00
TOTAL CYCLE: 01:08:60
CYCLE COMPLETE

06/15/2021 07:30 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____