

-----  
Surrey, BC  
Canada  
-----

STATIM 5000 S501R711  
SN 511018F00076  
Unit # : 000  
WATER QUALITY  
2.0uS / 1.2ppm  
CYCLE NUMBER 000038  
16:35 10/08/2020  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:24  
PRESSURIZING 2:37  
STERILIZING 3:45  
136.4 C 227kPa 3:45  
136.4 C 227kPa 4:15  
136.5 C 227kPa 4:45  
136.1 C 224kPa 5:15  
136.2 C 225kPa 5:45  
136.3 C 226kPa 6:15  
136.5 C 227kPa 6:45  
136.4 C 227kPa 7:15  
VENTING 7:15  
AIR DRYING 8:08  
CYCLE COMPLETE 10:08  
-----

Digital Signature #  
138266402A2319C8  
-----