
Midmark M11 - v1.0.0
Total Cycles: 4292

Name: RITTER 1
Sterilizer ID: V2178198
02/12/2024 10:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.0 F 0.0
2:00 162.0 F 1.5
4:00 171.5 F 4.8
6:00 193.6 F 9.0
8:00 215.1 F 8.8
10:00 227.3 F 7.6
12:00 236.4 F 12.4
14:00 236.4 F 9.1
16:00 240.2 F 10.2
18:00 246.1 F 13.1
20:00 257.8 F 19.3
22:00 268.4 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 272.6 F 28.9
2:00 271.6 F 28.4
3:00 272.8 F 29.1
4:00 272.0 F 28.4

Min 271.6 F 28.3
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.4 F 0.0
5:00 233.7 F 0.0
10:00 219.1 F 0.0
15:00 228.1 F 0.0
20:00 228.8 F 0.0
25:00 234.1 F 0.0
30:00 230.7 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 23:60
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:02:58
CYCLE COMPLETE

02/12/2024 11:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____