

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4102

Name: RITTER 1  
Sterilizer ID: V2178198  
12/11/2023 10:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 165.8 F 0.0  
2:00 160.8 F 1.3  
4:00 164.1 F 4.0  
6:00 202.0 F 8.0  
8:00 212.0 F 12.5  
10:00 219.1 F 5.6  
12:00 231.0 F 9.9  
14:00 235.6 F 9.4  
16:00 241.0 F 11.5  
18:00 238.4 F 9.7  
20:00 250.9 F 15.7  
22:00 261.5 F 21.8  
24:00 271.4 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.6 F 29.1  
2:00 272.4 F 29.1  
3:00 271.7 F 29.0  
4:00 271.4 F 28.7

Min 270.7 F 28.1  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.3 F 0.0  
5:00 197.1 F 0.0  
10:00 197.5 F 0.0  
15:00 196.7 F 0.0  
20:00 193.5 F 0.0  
25:00 191.3 F 0.0  
30:00 192.4 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 24:35  
STERILIZING: 4:00  
VENTING: 3:46  
DRYING: 30:00  
TOTAL CYCLE: 01:03:39  
CYCLE COMPLETE

12/11/2023 11:33 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_