
Midmark M11 - v1.0.5
Total Cycles: 2424

Name: RITTER2
Sterilizer ID: V2313228
05/29/2023 07:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.5 F 0.0
2:00 75.4 F 0.0
4:00 100.4 F 0.0
6:00 113.8 F 0.0
8:00 143.4 F 0.3
10:00 169.0 F 3.6
12:00 199.4 F 7.0
14:00 221.4 F 10.8
16:00 226.9 F 6.6
18:00 231.5 F 9.3
20:00 238.6 F 11.5
22:00 238.2 F 11.0
24:00 236.1 F 8.9
26:00 249.2 F 15.2
28:00 259.8 F 21.4
30:00 270.0 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 271.9 F 29.6
2:00 272.1 F 29.2
3:00 271.8 F 29.5
4:00 272.4 F 29.4

Min 271.2 F 28.9
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 212.1 F 0.0
10:00 211.0 F 0.0
15:00 225.8 F 0.0
20:00 221.5 F 0.0
25:00 221.0 F 0.0
30:00 218.5 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 30:47
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:08:44
CYCLE COMPLETE

05/29/2023 08:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____