
Midmark M11 - v1.0.0
Total Cycles: 3502

Name: RITTER 1
Sterilizer ID: V2178198
07/07/2023 10:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.4 F 0.0
2:00 196.6 F 3.8
4:00 213.1 F 11.6
6:00 226.9 F 7.6
8:00 236.7 F 9.5
10:00 244.1 F 11.9
12:00 247.4 F 13.7
14:00 249.1 F 14.6
16:00 264.8 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 271.9 F 28.0
2:00 272.5 F 28.3
3:00 272.9 F 28.7
4:00 272.7 F 28.9

Min 271.6 F 27.9
Max 273.0 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 223.5 F 0.0
5:00 230.1 F 0.0
10:00 212.7 F 0.0
Dry Cycle Aborted

FILLING: 1:34
HEATING: 17:33
STERILIZING: 4:00
VENTING: 2:55
DRYING: 10:09
TOTAL CYCLE: 00:36:11

07/07/2023 11:35 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____