
Midmark M11 - v1.0.0
Total Cycles: 4464

Name: RITTER 1
Sterilizer ID: V2178198
04/17/2024 08:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.6 F 0.0
2:00 70.1 F 0.2
4:00 81.3 F 1.1
6:00 114.9 F 3.3
8:00 141.9 F 6.0
10:00 166.5 F 9.6
12:00 194.6 F 13.2
14:00 215.9 F 4.1
16:00 222.8 F 7.2
18:00 226.2 F 10.7
20:00 228.1 F 7.3
22:00 237.6 F 9.9
24:00 238.0 F 9.2
26:00 245.2 F 12.7
28:00 256.7 F 18.7
30:00 267.0 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.5 F 28.5
2:00 271.8 F 28.4
3:00 272.2 F 28.4
4:00 271.8 F 28.4

Min 271.4 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.6 F 0.0
5:00 226.2 F 0.0
10:00 219.8 F 0.0
15:00 241.8 F 0.0
20:00 226.4 F 0.0
25:00 223.2 F 0.0
30:00 209.0 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:33
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:09:43
CYCLE COMPLETE

04/17/2024 09:20 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____