
Midmark M11 - v1.0.0
Total Cycles: 5346

Name: RITTER 1
Sterilizer ID: V2178198
02/19/2025 07:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.6 F 0.0
2:00 152.0 F 2.4
4:00 159.3 F 5.8
6:00 189.7 F 9.5
8:00 213.2 F 7.1
10:00 219.4 F 7.4
12:00 230.4 F 11.9
14:00 231.1 F 8.0
16:00 243.4 F 13.5
18:00 242.1 F 11.6
20:00 254.3 F 17.8
22:00 265.1 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.4 F 29.2
2:00 271.2 F 28.5
3:00 272.2 F 29.4
4:00 271.8 F 28.8

Min 271.1 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 202.2 F 0.0
10:00 208.3 F 0.0
15:00 213.1 F 0.0
20:00 215.6 F 0.0
25:00 222.0 F 0.0
30:00 223.4 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 23:50
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:02:36
CYCLE COMPLETE

02/19/2025 08:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____