

-----  
Midmark M11 - v1.0.0  
Total Cycles: 740

Name: RITTER 1  
Sterilizer ID: V2178198  
03/24/2021 09:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 95.2 F 0.0  
2:00 111.7 F 0.0  
4:00 118.9 F 0.0  
6:00 122.0 F 0.2  
8:00 162.0 F 2.9  
10:00 194.5 F 6.7  
12:00 224.7 F 10.8  
14:00 230.5 F 6.4  
16:00 236.2 F 10.9  
18:00 233.6 F 7.7  
20:00 243.0 F 11.5  
22:00 245.5 F 12.6  
24:00 258.0 F 19.3  
26:00 269.3 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 27.9  
1:00 272.3 F 28.3  
2:00 271.6 F 28.6  
3:00 272.2 F 28.8  
4:00 272.3 F 29.0

Min 271.0 F 27.8  
Max 273.1 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.1 F 0.0  
5:00 226.2 F 0.0  
10:00 236.7 F 0.0  
15:00 239.1 F 0.0  
20:00 229.4 F 0.0  
25:00 230.2 F 0.0  
30:00 224.5 F 0.0

Drying Complete

FILLING: 0:51  
HEATING: 26:60  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 01:04:51  
CYCLE COMPLETE

03/24/2021 10:28 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01