
Midmark M11 - v1.0.0
Total Cycles: 2259

Name: RITTER 1
Sterilizer ID: V2178198
08/29/2022 01:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.5 F 0.0
2:00 193.8 F 0.5
4:00 199.5 F 2.1
6:00 220.0 F 6.8
8:00 233.5 F 11.7
10:00 231.8 F 8.2
12:00 244.2 F 11.5
14:00 244.2 F 12.7
16:00 243.4 F 12.0
18:00 257.2 F 19.0
20:00 268.2 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.7 F 28.7
2:00 271.6 F 28.2
3:00 272.7 F 28.7
4:00 271.6 F 28.0

Min 271.5 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 216.3 F 0.0
10:00 235.9 F 0.0
15:00 234.4 F 0.0
20:00 226.4 F 0.0
25:00 224.1 F 0.0
30:00 222.9 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 21:12
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:41
CYCLE COMPLETE

08/29/2022 02:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____