

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3704

Name: RITTER2  
Sterilizer ID: V2313228  
06/17/2024 08:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.5 F 0.0  
2:00 72.8 F 0.0  
4:00 101.2 F 0.4  
6:00 108.0 F 2.1  
8:00 139.2 F 4.5  
10:00 162.7 F 7.4  
12:00 181.5 F 10.8  
14:00 204.0 F 4.6  
16:00 209.7 F 6.7  
18:00 220.8 F 10.4  
20:00 229.9 F 8.9  
22:00 235.6 F 10.5  
24:00 235.0 F 8.4  
26:00 246.4 F 14.0  
28:00 257.3 F 19.9  
30:00 267.5 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.1  
1:00 271.8 F 29.7  
2:00 272.1 F 29.3  
3:00 271.4 F 29.4  
4:00 272.5 F 29.7

Min 271.1 F 28.9  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.4 F 0.0  
5:00 210.1 F 0.0  
10:00 190.2 F 0.0  
15:00 185.5 F 0.0  
20:00 183.3 F 0.0  
25:00 183.0 F 0.0  
30:00 199.2 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 31:20  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:09:28  
CYCLE COMPLETE

06/17/2024 10:00 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_