

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3436

Name: RITTER 1  
Sterilizer ID: V2178198  
06/23/2023 12:06 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 214.8 F 0.0  
2:00 200.8 F 3.7  
4:00 212.6 F 8.8  
6:00 224.4 F 8.4  
8:00 232.0 F 9.9  
10:00 235.0 F 8.3  
12:00 245.1 F 13.5  
14:00 244.4 F 12.4  
16:00 257.2 F 19.2  
18:00 268.6 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.6 F 29.1  
2:00 271.9 F 28.4  
3:00 271.9 F 28.7  
4:00 272.6 F 29.1  
  
Min 271.4 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.3 F 0.0  
5:00 229.3 F 0.0  
10:00 228.8 F 0.0  
15:00 228.9 F 0.0  
20:00 230.3 F 0.0  
25:00 229.8 F 0.0  
30:00 229.8 F 0.0  
Drying Complete

FILLING: 2:07  
HEATING: 19:14  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:37  
CYCLE COMPLETE

06/23/2023 01:04 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_