

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3912

Name: RITTER2  
Sterilizer ID: V2313228  
09/06/2024 11:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.6 F 0.0  
2:00 180.3 F 2.6  
4:00 192.6 F 7.5  
6:00 214.1 F 12.9  
8:00 224.6 F 8.6  
10:00 235.3 F 10.5  
12:00 241.4 F 12.9  
14:00 242.4 F 12.3  
16:00 255.8 F 19.3  
18:00 267.5 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.4  
1:00 272.3 F 29.8  
2:00 272.0 F 30.0  
3:00 272.2 F 29.7  
4:00 272.0 F 30.1  
  
Min 271.4 F 29.3  
Max 272.7 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.5 F 0.0  
5:00 214.8 F 0.0  
10:00 217.6 F 0.0  
15:00 213.4 F 0.0  
20:00 207.6 F 0.0  
25:00 200.5 F 0.0  
30:00 201.5 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 19:19  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 00:58:22  
CYCLE COMPLETE

09/06/2024 12:32 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_