
Midmark M11 - v1.0.5
Total Cycles: 4480

Name: RITTER2
Sterilizer ID: V2313228
04/02/2025 04:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.9 F 0.0
2:00 139.2 F 1.5
4:00 143.8 F 4.5
6:00 161.2 F 8.4
8:00 185.5 F 12.7
10:00 212.3 F 5.8
12:00 227.2 F 10.1
14:00 233.3 F 9.0
16:00 238.0 F 11.2
18:00 236.2 F 9.2
20:00 248.5 F 15.2
22:00 259.1 F 21.1
24:00 268.7 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.4
1:00 272.7 F 30.2
2:00 271.9 F 30.4
3:00 271.7 F 29.8
4:00 272.6 F 30.2

Min 270.9 F 29.4
Max 273.2 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.7 F 0.0
5:00 179.1 F 0.0
10:00 196.1 F 0.0
15:00 198.0 F 0.0
20:00 192.6 F 0.0
25:00 190.6 F 0.0
30:00 192.6 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 25:11
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:04:02
CYCLE COMPLETE

04/02/2025 05:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____