
Midmark M11 - v1.0.0
Total Cycles: 1146

Name: RITTER 1
Sterilizer ID: V2178198
09/08/2021 08:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.0 F 0.0
2:00 155.8 F 1.1
4:00 172.3 F 4.0
6:00 186.6 F 8.1
8:00 206.0 F 12.9
10:00 218.7 F 6.2
12:00 232.3 F 11.0
14:00 231.4 F 7.5
16:00 244.1 F 13.4
18:00 242.8 F 11.9
20:00 255.3 F 18.3
22:00 266.3 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 272.9 F 29.2
2:00 271.8 F 28.4
3:00 272.8 F 29.3
4:00 272.3 F 28.8

Min 271.6 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 204.6 F 0.0
10:00 212.4 F 0.0
15:00 212.4 F 0.0
20:00 212.5 F 0.0
25:00 215.5 F 0.0
30:00 216.3 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 23:39
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:01:51
CYCLE COMPLETE

09/08/2021 09:14 AM
ACCEPTED: X
REJECTED:
Operator: U01